



**Registration No. CNAS L0641**

Tianjin Institute of Metrological Supervision and Testing

**Verification Certificate**

Certificate No: 21100828

Name of Client BASS INSTRUMENTS

Name of Instrument ELECTROMAGNETIC FLOWMETER

Model/Type EMDE.RM.0100.16.02.04.AC.04.01.NN.CS.02

Serial No 21100828

Manufacturer TIANJIN SURE INSTRUMENT CO.LTD

Verification Regulation JJG1033-2007 , ISO 8316:1995

Conclusion 0.2 %

Approved by \_\_\_\_\_

Checked by \_\_\_\_\_

Verified by \_\_\_\_\_



Date of Verification 2021 Year 11 Month 16 Day

Date of Expiry 2023 Year 11 Month 15 Day

**Accredited in accordance with ISO/IEC 17025:2017**





### Registration No. CNAS L0641

Verification Certificate of TIMST

Certificate No 21100828

Tianjin Institute of Metrological Supervision and Testing is National Legal Metrological Verification Organization

The number of Certificate of metrological authorization to The Legal Metrological Verification Institution is [2017]01009

#### Verification Environmental Conditions and Location

Location	TIANJIN				
Temperature	5°C	Relative Humidity	42 %	Others	/

#### Measurement Standart for Verification

Name	Measuremet Range	Uncertainty/Accuracy/ Maximum Permissible Error	The Number of Measurement Standart Certificate	Date of Expiry
Standart Water Flow Meter	0.009...1000 m <sup>3</sup> /h	U <sub>rel</sub> =0.16% (k=2)		25.06.2022

#### Main Measuring Instruments Used This Verification

Name	Serial No	Measuremet Range	Uncertainty/Accuracy/ Maximum Permissible Error	Certificate No	Date of Expiry
Calibration Rig	201211	DN(2.5...500)mm	U <sub>rel</sub> =0.15 (k=2)	FLLdr19084562	20.09.2023

The above measuring value of instruments could be traced to primary standart of People's Republic of China



Our Institute only takes responsibility for complete certificates stamped with specified verification seal of 'Tianjin Institute of Metrological Supervision and Testing'. The results represent only the sample(s) of this verification.

### Accredited in accordance with ISO/IEC 17025:2017